

**Certificate of Analysis**

DVC-Rev:	10674-02
Rev By:	JLA
Check By:	BT

**Product Name / Specification:** 2M Sodium Acetate, pH 4.5. 500mL, Sterile.

**Version:**  
6.0

**Part Number:** S0292  
**LotNumber:** S029223C2101  
**Date Manufactured:** 3/22/2021  
**Date of Re-assay:** 3/22/2024

**Date QC Released:** 3/25/2021

**Released By:**



John Brusven, QC Manager

**Appearance:** Clear, colorless liquid  
**Method of Sterilization:** 0.2µm Sterile filtration  
**Storage Conditions:** Room temperature, 2-30°C.

<u>QA Testing:</u>	<u>Result:</u>	<u>Expected:</u>
Conductivity	37.45mS/cm	N/A
Density	1.04008g/cm³	N/A
pH	4.50	4.50 ± 0.04
Sterility	Pass	Pass